



## Legislation Text

File #: 20-0559, Version: 1

MOTION TO APPROVE THE PURCHASE OF SEVENTY (70) CITY- STANDARD LIFEPAK CR2 DEFIBRILLATORS AND FIVE (5) LIFEPAK CR2 TRAINERS FROM THE SOLE SOURCE DISTRIBUTOR, STRYKER CORPORATION, IN AN AMOUNT NOT TO EXCEED \$140,570.75, UTILIZING PRICING ESTABLISHED BY THE NATIONAL ASSOCIATION OF STATE PROCUREMENT OFFICERS (NASPO) CONTRACT # NASPO17 #OK-SW-300, PURSUANT TO SECTIONS 35.18(C)(3) AND 35.18(C)(5) OF THE CITY'S PROCUREMENT CODE.

### PROCUREMENT PROCESS TAKEN:

- ***Chapter 35 of the City's Code of Ordinances is titled "PROCUREMENT PROCEDURES, PUBLIC FUNDS."***
- ***Section 35.18 of the City's Code of Ordinances is regarding "COMPETITIVE BIDDING OR COMPETITIVE PROPOSALS REQUIRED; EXCEPTIONS."***
- ***Section 35.18(C) states that "Only the following situations are exempt from the competitive bid and competitive proposal requirements of this section:"***
- ***Section 35.18(C)(3) states that "City standard, single-source and sole- source commodities or services. City standard, single- source and sole-source commodities or services are exempt from this section."***
- ***Section 35.18(C)(5) states that "Commodities or services that are the subject of contracts with the state its political subdivisions or other governmental entities including the United States Government, are exempt from the competitive procurement process."***
- ***Section 35.21(A)(1) states, "An initial purchase of, or contract for, commodities or services, in excess of \$25,000, shall require the approval of the City Commission, regardless of whether the competitive bidding or competitive proposal procedures were followed."***

### SUMMARY EXPLANATION AND BACKGROUND:

1. On August 7, 2019, the City Commission approved the LifePak Cardiac Monitor / Defibrillator from Stryker Corporation as a City Standard.
2. The LIFEPAK CR2 is the most advanced defibrillator on the market, and its defibrillation pads are compatible with our LIFEPAK 15 monitors. This feature allows our responding personnel to quickly disconnect the CR2 defibrillator and connect our LIFEPAK 15 monitor without having to remove the defibrillation pads commonly found with different defibrillator manufacturers, thus saving valuable time.

**3. Additional highlights:**

- Performs rhythm analysis without the need to pause chest compressions. This is the only AED (Automated External Defibrillator) on the market that currently has this function. Continuous chest compressions with minimal interruptions is an AHA standard.
- Child and Adult modes using the same defibrillation pads.
- Fully automated, eliminating guess work.
- LIFELINK AED Program Manager is included. Via Wi-Fi, we are able to monitor each AED location as well as battery and pad life conditions.
- LIFEPAK CR2 Trainer provides realistic training in the use of the LIFEPAK CR2 defibrillator.

4. Stryker Corporation, the sole source distributor for LIFEPAK, has provided a quote for the purchase and shipping of seventy (70) LIFEPAK CR2 Defibrillators and five (5) LIFEPAK CR2 Trainers, utilizing pricing established from the National Association of State Procurement Officers (NASPO) contract # NASPO17 #OK-SW-300.

Description	Qty	Cost	Total
LIFEPAK CR2 Defibrillator	70	\$1,968.75	\$137,812.50
LIFEPAK CR2 Trainer	5	\$551.65	\$2,758.25
	<b>Total</b>		<b>\$ 140,570.75</b>

5. Request the City Commission to approve the purchase of seventy (70) City- Standard LIFEPAK CR2 Defibrillators and five (5) LIFEPAK CR2 Trainers from the sole source distributor, Stryker Corporation, in an amount not to exceed \$140,570.75, utilizing pricing established by the National Association of State Procurement Officers (NASPO) contract # NASPO17 #OK-SW-300, pursuant to Sections 35.18(C)(3) and 35.18(C)(5) of the City's Procurement Code.

**FINANCIAL IMPACT DETAIL:**

- a) **Initial Cost:** \$140,570.75
- b) **Amount budgeted for this item in Account No:** Funds are available in the # 1-529-4003-64352 (Special Equipment - Rescue)
- c) **Source of funding for difference, if not fully budgeted:** Not Applicable.
- d) **5 year projection of the operational cost of the project** Not Applicable.
- e) **Detail of additional staff requirements:** Not Applicable.